

How often should CFEZ be standardized

The standardization interval for the CFEZ has been semi-automated. The CFEZ will require re-standardization if one of the following occurs, A. 8 hours elapses since last standardization. B. Internal temperature of sensor changes by 5 degrees Celsius.